


# CBC-CAL


## HEMATOLOGY CALIBRATOR

### Reference Values provided by DYMIND

### CALIBRATOR

 PLUS0425

 2025-03-07

 2025-05-05

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.20
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	132
	MCV	fL	92.3
	PLT	$\times 10^9/L$	237
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.99
	RBC	$\times 10^{12}/L$	4.48
	HGB	g/L	128
	MCV	fL	84.2
	PLT	$\times 10^9/L$	236
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.39
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	132
	MCV	fL	89.7
	PLT	$\times 10^9/L$	243
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	8.96
	RBC	$\times 10^{12}/L$	4.46
	HGB	g/L	125
	MCV	fL	91.7
	PLT	$\times 10^9/L$	228

#### 【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


# CBC-CAL


## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**

 PLUS0425

 2025-03-07

 2025-05-05

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.15
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	127
	MCV	fL	87.6
	PLT	$\times 10^9/L$	233
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.87
	RBC	$\times 10^{12}/L$	4.46
	HGB	g/L	135
	MCV	fL	92.6
	PLT	$\times 10^9/L$	228

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.