

# **XE-RET**

## **HEMATOLOGY CONTROL**

### Reference Values provided by DYMIND

CONTROL

LOT 35031

M 2025-02-05

	2023 03								
Applicable Instruments	Parameter	Unit	LOT 35031001	LOT 35031002	LOT 35031003				
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	×10 <sup>12</sup> /L	0.0329 ±0.0165	0.2121 ±0.1273	0.3179 ±0.1200				
	RET%	%	0.59 ±0.30	4.73 ±1.89	13.25 ±4.50				
	LFR	%	89.0 ±27.0	89.8 ±27.0	89.6 ±27.0				
	MFR	%	8.5 ±13.0	8.3 ±12.0	8.4 ±13.0				
	HFR	%	2.5 ±4.0	1.9 ±3.0	2.0 ±3.0				
	IRF	%	11.0 ±17.0	10.2 ±15.0	10.4 ±16.0				
	RHE	pg	17.8 ±7.0	14.7 ±6.0	14.5 ±6.0				

#### [NOTE]

- 1. The controls should be stored in refrigerator (2 ℃~8 ℃). After opening, it will keep stable for 14 days when it is stored airtight at 2 ℃~8 ℃.
- 2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature ( $15\%\sim30\%$ ).
- 3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- 4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.







35031001

35031002

35031003



# **XE-RET**

## **HEMATOLOGY CONTROL**

### Reference Values provided by DYMIND

CONTROL

LOT 35031

M 2025-02-05

2025-05-05

Applicable Instruments	Parameter	Unit	35031001		35031002		35031003	
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	×10 <sup>12</sup> /L	0.0140	±0.0112	0.1634	±0.0817	0.1762	±0.0881
	RET%	%	0.26	±0.21	3.78	±1.89	7.40	±3.70

#### [NOTE]

- 1. The controls should be stored in refrigerator (2 ℃~8 ℃). After opening, it will keep stable for 14 days when it is stored airtight at 2 ℃~8 ℃.
- 2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature ( $15^{\circ}\text{C} \sim 30^{\circ}\text{C}$ ).
- 3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- 4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.







35031001 35031002 35031003